

measured would be estimated to be a weighted average of its component properties. Upon studying the Examples in the specification, it will be noted that the bending rigidity of the heterogeneous braids in Examples 1 and 2 do not follow the Rule of Mixtures, but surprisingly show an enhanced bending rigidity relative to the weighted average of their filament components. This behavior is not achieved when dissimilar individual filaments are randomly braided to form the braided suture.

In setting forth the claimed invention, the heterogeneous braid does not encompass braided sutures with randomly braided individual filaments, as described in detail above. Further, the claimed heterogeneous braid could not be construed to cover known braids which have a core of longitudinally extending yarns composed of filaments of a first fiber-forming material, and a sheath of braided yarns composed of a second set of filaments of a dissimilar fiber-forming material. This braid construction does not fall within the scope of the claimed braid because these sheath yarns are not in direct intertwining contact with any of the core yarns. In other words, none of the sheath yarns are braided about a core yarn, but simply shroud the core yarns to form the sheath construction.

#### Analysis of the Rejection

1. Claims 21 and 23 were rejected under 35 USC §102(b) as being clearly anticipated by Doddi et al. ("Doddi"). Doddi does not anticipate the claimed suture, and therefore this rejection should be withdrawn.

The Examiner has correctly pointed out that Doddi does indeed disclose a surgical suture comprising filaments of two different polymers in a braided configuration (column 9, lines 47-56).

However, as discussed in detail above, more is required to meet the limitations of the claimed suture than just a disclosure concerning filaments of two different polymers in a braided configuration. Doddi teaches nothing more than braiding individual filaments, and fails to provide any guidance as to how that braiding should be carried out. Therefore, one skilled in the art would be lead to believe that what Doddi had in mind was to simply braid individual filaments in a randomized fashion to fabricate a multifilament suture. It is important enough, however, to reemphasize again that the claimed braid requires the bundling of individual filaments into an integrated unit to form a multifilament yarn. It is this multifilament yarn which directly intertwines with another multifilament yarn to form Applicants' braid construction.

Since Doddi only teaches randomly braiding filaments of dissimilar fiber-forming materials, it does not anticipate the claimed braided suture. Doddi simply fails to enable one skilled in the art to construct a braided suture in the manner set forth by Applicants, and it is axiomatic that a reference which lacks enablement is deficient as a reference to anticipate a claimed invention. Accordingly, it is respectfully requested that the rejection of claims 21 and 23 under 35 USC §102(b) as being clearly anticipated by Doddi be withdrawn.

2. Claims 22 and 24 were rejected under 35 USC §103 as being unpatentable over Kaplan et al. ("Kaplan") taken with Doddi. The Examiner asserts it would have been obvious to substitute PET and PTFE fibers of Doddi for the filaments of Kaplan to arrive at Applicants' claimed suture. Applicants respectfully traverse this rejection for the reasons given below.

The Examiner correctly points out that Kaplan discloses a ligament prosthesis made from a core component and a braided sheath component as illustrated in Figures 3 and 4, and discussed at column 8, line 65, through column 9, line 34. However, Kaplan suffers from the same deficiencies as does Doddi, and therefore fails to teach or suggest the claimed braided suture.

Firstly, the Examiner has made specific reference to the Kaplan specification regarding the makeup of the core components and the sheath yarn component. The only component which has a braided construction is the sheath yarn component. It is clear from Figure 3 of Kaplan that none of the sheath yarn components are in direct intertwining contact with the core component. In other words, the sheath yarn component is a true "sheath" which shrouds the core but is not in any way integrally braided with the core. Therefore, since the core is not in a braided construction, its composition is irrelevant with respect to the claimed braided suture.

When the focus is shifted to the more relevant aspect of the Kaplan disclosure, specifically the sheath yarn component, the Examiner has correctly pointed out that the sheath yarn component may be "fabricated from individual filaments having more than two different chemical compositions, one or more of which optionally being non-absorbable". (Column 9, lines 25-28). However, Kaplan neither teaches nor suggests how his sheath yarn component is to be fabricated from these dissimilar individual filaments, nor is there any guidance to one skilled in the art as to how such dissimilar individual filaments are to be braided. Accordingly, just as was the case with the deficient Doddi reference, one skilled in the art could only be lead to randomly braid the dissimilar individual filaments into a braid construction.

The teaching of Kaplan once again lacks the essence of the claimed invention, which is: bundled filaments of a first fiber-forming material form a first set of a multifilament yarns, and at least one of these multifilament yarns is intertwined with a multifilament yarn composed of bundled filaments of a second fiber-forming material. To put it bluntly, Kaplan teaches randomized braiding, and the claimed suture sets forth a structured braid. This difference is not trivial, as pointed out with reference to the discussion of Applicant's specification, and particularly Examples 1 and 2.

It should also be pointed out here that even if Doddi and Kaplan were combined, their combined teachings would still fail to meet the limitations of the claimed braided suture. This is so because neither reference, taken singularly or in combination, discloses a structured braid set forth in the claims, but merely sets forth randomized braiding of individual filaments.

For all of the reasons given above, especially taken in light of the detailed discussion of the claimed braided suture and its surprising advantages, the rejection of claims 22 and 24 under 35 USC §103 as being unpatentable over Kaplan taken with Doddi is improper. Accordingly, it is respectfully requested that this rejection be withdrawn.


3. Applicants acknowledge with gratitude the withdrawal of the rejection of claims 21-24 under 35 USC §103 as being unpatentable over Burgess, expressed in the previous Office Action dated July 8, 1992. (Paper No. 3). It is presumed that Applicants' response to this rejection in their Amendment dated August 6, 1992, spelling out the distinctions between Burgess and the claimed

invention, clearly convinced the Examiner that the claimed surgical suture is patentable over this art.

4. The prior art made of record and not relied upon by the Examiner is duly noted, and does not affect the patentability of Applicants' claimed invention.

5. Since all formal requirements appear to have been met, and the claimed invention is patentable over the art of record or any other art of which Applicants are aware, Applicants respectfully solicit a Notice of Allowance at the Examiner's earliest convenience.

Respectfully submitted,

  
Matthew S. Goodwin  
Attorney for Applicant  
Reg. No. 32,839

Johnson & Johnson  
One Johnson & Johnson Plaza  
New Brunswick, NJ 08933-7003  
(908) 524-2794  
December 2, 1992